CardioChek® PA - Serial No:				Site Name:			
Daily Instrument Check Strip Test (and Fault) Log sheet – Use this section to record any faults that arise with the CardioChek® PA device when testing with the GREY strip NB: For each CardioChek® PA device you must use a separate log sheet							
Date	Pass (P)/ Fail (F)	Action taken to resolve		**Problem** (if check-strip test FAILS, what caused the failure?)	Further action (if problem has not resolved)	Date resolved	Operator initials
01/01/2015	*F*	*Clean cotton-bud "wet by alcohol wipe", dried with other end. Retested check-strip.*	*P*	*dirty optical glass*	*N/A*	*01/01/2015*	*JAC*
02/01/2015	*P*	*N/A*	*N/A*	*none*	*N/A*		*JAC*